

INCIDENT REPORTING IN EU MEMBER STATES

If any serious incident in relation to this device has occurred, please report without delay to the manufacturer and competent authority of your Member State.

SHIPPING DAMAGE

Please notify your distributor, if this product was received damaged.

SYMBOLS

BÜHLMANN use symbols and signs listed and described in ISO 15223-

1. For definition of symbols see the symbol glossary at:

www.buhmannlabs.ch/support/downloads/

In addition the following symbols and signs are used:



EN: electronic instruction for use available in different languages at:/ **BG:** електронни инструкции за употреба на различни езици на адрес:/ **CS:** elektronický návod k použití dostupný v různých jazycích na adrese:/ **DA:** elektronisk brugsanvisning på forskellige sprog på:/ **DE:** elektronische Gebrauchsanweisung in verschiedenen Sprachen verfügbar

unter:/ **EL:** ηλεκτρονικές οδηγίες χρήσης διαθέσιμες σε διάφορες γλώσσες στη διεύθυνση:/ **ES:** instrucciones de uso electrónicas disponibles en diferentes idiomas en:/ **ET:** elektrooniline kasutusjuhend, mis on saadaval erinevates keeltes aadressil:/ **FR:** un mode d'emploi électronique disponible en différentes langues à l'adresse:/ **HU:** külfönbőző nyelveken elérhető elektronikus használati utasítás a következő címen:/ **IT:** istruzioni elettroniche per l'uso disponibili in diverse lingue su:/ **LT:** elektroninės naudojimo instrukcijos įvairiomis kalbomis:/ **LV:** dažādās valodās pieejama elektroniska lietošanas instrukcija:/ **NO:** elektronisk instruksjon for bruk tilgjengelig på forskjellige språk på:/ **PL:** elektroniczna instrukcja obsługi dostępna w różnych językach na stronie:/ **PT:** instrução electrónica para utilização disponível em diferentes línguas em:/ **RO:** instrucțiuni electronice de utilizare disponibile în diferite limbi la adresa:/ **SK:** elektronický návod na použitie dostupný v rôznych jazykoch na:/ **SL:** elektronska navodila za uporabo so na voljo v različnih jezikih na:/ **SR:** elektronsko uputstvo za upotrebu dostupno na različitim jezicima na:/ **SV:** elektronisk bruksanvisning på olika språk på följande adress:

www.buhmannlabs.ch/support/downloads/

US Distribution

BÜHLMANN Diagnostics Corp

Amherst, NH 03031, USA

Tel: (844)300-979

info@buhmannlabs.com



BÜHLMANN fCAL® turbo

Calprotectin turbidimetric assay
for professional use

Calibrator Kit

B-KCAL-CASET

Version A4.2

For *In Vitro* Diagnostic Use

CLIA Complexity: High

Rx Only



Manufacturer

BÜHLMANN Laboratories AG

Baselstrasse 55

4124 Schönenbuch

Switzerland

Tel.: +41 61 487 1212

Fax: +41 61 487 1234

info@buhmannlabs.ch



[www.buhmannlabs.ch/
support/downloads/](http://www.buhmannlabs.ch/support/downloads/)

INTENDED USE

The BÜHLMANN fCAL® turbo Calibrator Kit is intended for use with the BÜHLMANN fCAL® turbo Reagent Kit for the determination of fecal calprotectin levels in extracted stool samples. Each calibrator establishes a point of reference for the working curve that is used to calculate test results from patient samples.

For laboratory use only.

CALIBRATOR VALUE

Calibrator values are assigned according to a value transfer protocol (ref. 1-2) and are indicated in the enclosed QC-data sheet. The calibrator material comprises blood-derived human calprotectin and is standardized against internal reference material.

REAGENTS SUPPLIED

Reagents	Quantity	Code	Preparation
Calibrators Calibrators 1-6 containing an assigned concentration of human calprotectin	1 x 6 vials 1 mL/vial	B-KCAL-CASET	Ready to use

Table 1

REAGENT STORAGE AND STABILITY

Unopened calibrators Store at 2-8 °C. Do not use kit past expiration date printed on the labels.
Opened calibrators Store for up to 3 months at 2-8 °C, capped.
Calibration curve stability Refer to the instrument specific application note.

Table 2

MATERIALS REQUIRED BUT NOT PROVIDED

Reagents	Quantity	Code
BÜHLMANN fCAL® turbo Reagent Kit Reaction Buffer (R1) Immunoparticles (R2)	1 vial/35 mL 1 vial/7 mL	B-KCAL-RSET B-KCAL-RSET-US ¹
BÜHLMANN fCAL® turbo Control Kit Controls low and high	3 x 2 vials 1 mL/vial	B-KCAL-CONSET

Table 3

¹ For US customers only

WARNINGS AND PRECAUTIONS

- This test is for *in vitro* diagnostic use only.
- This kit contains components classified in accordance with the Regulation (EC) No. 1272/2008: 2-methyl-4-isothiazolin-3-one hydrochloride (conc. ≥ 0.0015%), thus the reagents may cause allergic skin reactions (H317).

- Before measuring please equilibrate reagents, controls, calibrators and samples as described in the application note.
- Do not mix calibrators of different lots or switch caps between reagents.
- Avoid evaporation of the calibrator.
- The calibrator contains components of human origin. Although tested and found negative for HBV, HCV and HIV, the calibrators should be handled as if capable of transmitting infections and should be handled in accordance with Good Laboratory Practices (GLP) using appropriate precautions. Disposal of any discarded materials should be in accordance with local requirements.

ASSAY PROCEDURE

Application notes / assay Installation

The assay procedure for the BÜHLMANN fCAL® turbo has been established on several clinical chemistry analyzers. Validated application notes describing installation and analysis on specific instruments are available from BÜHLMANN upon request.

Establishment of the calibration curve

The BÜHLMANN fCAL® turbo Calibrator kit is used to establish a six point calibration curve according to the instrument manual. Calibrator values are lot-specific. A new calibration must be performed for each new calibrator and reagent lot. Otherwise, calibration should be performed every one to two months according to the instrument specific application notes. Refer to the enclosed QC-data sheet for assigned calibrator values. Contact BÜHLMANN support if calibration cannot be performed without error.

QC controls

The calibration curve must be validated with controls, low and high, each day before running patient fecal sample extracts. Refer to the instruction for use for BÜHLMANN fCAL® turbo Control kit for further information.

CHANGELOG

Date/Version	Change
2023-07-12 / A4.2	Revision of chapter <i>Symbols</i>

REFERENCES

1. Blirup-Jensen et al.: Clin Chem Lab Med 2001; 39, 1110 – 22.
2. Blirup-Jensen et al.: Clin Chem Lab Med 2008; 46, 1470 – 9.